



Certificate of Compliance

Certificate: 1188789

Master Contract: 204767

Project: 1800498

Date Issued: 2006/08/23

Issued to: Camille Bauer AG
Aargauerstrasse 7
Wohlen, 5610
Switzerland
Attention: Mr. Markus Fahndrich

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Eknath Desai; C.E.T.

Authorized by: Renzo Pupulin, C.E.T., Product Group Manager

PRODUCTS

CLASS 3631 84 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

CLASS 3631 04 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

Programmable Multi-Transducer, Model SINEAX M 5xx, fixed equipment, rated 24-60V dc/ac, 50/60Hz, 7VA or 85-230V dc/ac, 50/60Hz, 7VA, or powered from measuring inputs, Equipment Class II, Pollution Degree 2, Installation Category III.

Where x in the model designation may be any number 0 to 9 or letter A to Z or blank denoting SELV / ELVEL secondary circuits or minor mechanical differences.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



CSA INTERNATIONAL

Certificate: 1188789

Master Contract: 204767

Project: 1800498

Date Issued: 2006/08/23

APPLICABLE REQUIREMENTS

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No 1010.1-92 -Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1).

CAN / CSA C22.2 No 1010.1B 97 - Amendment 2 to CAN / CSA C22.2 No. 1010.1 92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements".

UL Std No 3111-1 - Equipment for Measurement Use; Part I: General Requirements.



Supplement to Certificate of Compliance

Certificate: 1188789

Master Contract: 204767

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1800498	2006/08/23	UPDATE REPORT # 1188789 TO COVER MINOR ALTERNATE CONSTRUCTION.

History

1188789 - August 17, 2001 - Original certification.

Supplement Notes

1285712 - January 27, 2002 - Update to cover an alternate construction.

1313255 - April 26, 2002 - Update to cover an alternate construction.